

D 92 - ISO NF EN 22592 - ISO 2592 - IP 36 - DIN 51376

SCOPE

Standard Test Method for Flash and Fire Points by Cleveland Open Cup Tester

**MANUAL CLEVELAND FLASH POINT
- MODEL NCL 120 -
- REF 942611 -**

MAIN CHARACTERISTICS

- ✓ Electric heater
- ✓ Manual heating slope regulation
- ✓ Manual temperature reading
- ✓ Manual pilot flame

SCOPE OF DELIVERY

NCL 120 delivered ready for use with :

REF 941609 Gas tubing, dia. 2x4mm (1 m)

REF 941613 Cleveland cup with handle

REF 941615 Stoptherm insulating board

REF 11498 ASTM thermometer (11 C)

OPTIONAL ACCESSORIES

REF 9417444 Pair of safety gloves

REF 40484 Protective screen

CONSUMMABLES

REF 941609 1m of gas tubing



ORDERING INFORMATION

REF 942611

NCL 120 For use on AC 230 V, 50 Hz - 0.3 A

REF 962611 / 115V

NCL 120 For use on AC 115V, 50/60Hz

CONTACT : sales@normalab.com

NORMALAB FRANCE SAS
ZA Caux Multipôles 1 - F-76 190 Valliquerville
Tel. : +33 232 700 100
Fax : +33 232 704 732

www.normalab.com

NCL 120 Leaflet 15

DISTRIBUTED BY

G-Labo Germany
Bgm.-Horneffstr.26
69509 Mörlenbach
Tel.: + 49 6209 797100
Fax: + 49 6209 797101
Mail: info@g-labo.de
Web: www.g-labo.de