

IP 170, ISO 13736, NFT 66009 & DIN 51755**SCOPE**

Petroleum products and other liquids – Determination of flash point-Abel closed cup method.

**MANUAL ABEL FLASH POINT APPARATUS,
WITH ELECTRIC HEATER AND GAS TEST JET
- MODEL **NAB 110** -
- REF 941601 -****MAIN CHARACTERISTICS**

- ✓ A traditional copper water bath
- ✓ Oil cup
- ✓ Gas test jet
- ✓ Manual stirrer
- ✓ Thermometers
- ✓ Electric heater

SCOPE OF DELIVERY

NAB 110 delivered ready for use with :

- REF 941604** Cup
- REF 9416041** Lid shutter and stirrer with gas test jet
- REF 21227** IP thermometer (74 C)
- REF 21228** IP thermometer (75 C)
- REF 11493** ASTM Thermometer (6C)

OPTIONAL ACCESSORIES**Cryostat for test below 40°C**

- REF 23203** TLC 30 (-30 to +60°C ±0.05)
- REF 23204** TLC 40 (-40 to + 20 °C ±0.05)
- REF 941143** Energy regulator
- REF 941597** Dry ice machine
- REF 21229** IP thermometer (2 C)

**ORDERING INFORMATION****REF 941601**

NAB 110 For use on AC 230V, 50 Hz, 2A

(W)250x (D) 280x (H) 420 mm (±7 kg)

CONTACT : sales@normalab.com

NORMALAB FRANCE SAS
14 rue des Lilas - F-76 210 Lintot
Tel. : +332 35 38 59 59
Fax : +332 35 38 78 55

www.normalab.com

NAB 110 Brochure 14

DISTRIBUTED BY

G-Labo Germany
Bgm.-Horneffstr.26
69509 Mörlenbach
Tel.: + 49 6209 797100
Fax: + 49 6209 797101
Mail: info@g-labo.de
Web: www.g-labo.de