

# NPM 131

## PENSKY - MARTENS FLASH POINT MANUAL TESTER



### STANDARDS

ASTM D 93, IP 34, ISO 2719, NF M07019, DIN 51758, GBT 261, JIS K 2265-3 and related methods.

### SCOPE

These test methods cover the determination of the flash point of petroleum products in the temperature range from 40 to 370°C by a manual or an automated Pensky-Martens closed-cup apparatus. Procedure A is applicable to distillate fuels, lubricating oils, and other homogeneous petroleum liquids. Procedure B is applicable to residual fuel oils, cutback residua, used lubricating oils, mixtures of petroleum liquids.

